



NO_x long-term measurements at SIRTA station (Paris region, France)

Nicolas Bonnaire¹, Lorna Foliot¹, Valérie Gros¹, Marc-Antoine Drouin², Christophe Boitel²,
Aurélien Chauvigné³,

¹ Laboratoire des Sciences du Climat et de l'Environnement, LSCE, UMR CNRS-CEA-UVSQ, IPSL, Gif-sur-Yvette, France


² Laboratoire de Météorologie Dynamique, LMD/IPSL, Palaiseau, France

³ Centre de Données et Services AERIS/ICARE - UAR 2877, 59650 Villeneuve d'Ascq


SIRTA station: peri-urban station in Paris area



Regional background




Meteo & O₃ (2023)





- Dry winter
- Hot year but not scorching heat events

O₃ less dependent of wind direction
20 ppb < O₃ < 55 ppb

Meteo : Wind (2023)



- Higher mixing ratio : North-East wind
- Most frequently wind : South-west



- South-west : < 1 ppb
- North-East until 10 ppb

- South-west : < 4 ppb
- North-East until 20 ppb

Instrumentation

Instrumentation :

NO & NO₂ : API TELEDYNE T200UP, NO by ozone induced chemiluminescence detection (O₃-CLD) and measurement of NO₂ by photolytic conversion of NO₂ to NO followed by chemiluminescence detection (BLC-O₃-CLD)

maintenance / QA during measurement :

1/month : calibration with 30 ppb NO (dynamic dilution & GPT)

4/year : converter efficiency, near of 40 %

May 2023 : Lamp of BLC was changed and leak issue on the BLC

1/year linearity of analyzer

Scale : NPL # 1635, NO = 10,00 ± 0,06 ppm in N₂ (Since April 2014)

Every two week : filter change

Automatic download each day of parameters for Offline checks (chamber temp, pressure, flow, O₃-supply)


laboratory journal

Problem : leak on the UV lamp (O₃ generator)) in the calibrator (No excess NO)




2016-10-20_pm

scaled data (according to "MOHp2016_sensitivities.log")





2016-10-14_am



scaled data (according to "MOHp2016_sensitivities.log")



Yearly-Plots : NO & NO₂ , 2023 (daily mean)



Daily-Plots : NO & NO₂ , 2023 (hourly medians)



2023 Data information

- Level 0a & 1a submitted
- Level 0a : need raw counts and sensibility of NO ((pmol/mol)/cps) but it is impossible to get these parameters with Teledyne device
- The Ozone generated as part of the photolytic process back titrates the converted NO. This phenomenon is non-linear and leads to a situation where the calculated CE value is less at higher concentrations of NO₂. So in Teledyne device there are two converter efficiency (For SIRTA station at 16 and 70 ppb)
- Data need to be resubmitted (since Juin), to apply a new converter efficiency