



ACTRIS

CiGas

Implementation Status



Germany



France



Finland



Switzerland



Germany

NO_x/VOC QA workshop 2023; Online – April, 17th - 19th 2023



This project receives funding from the European Union's Horizon 2020 research and innovation programme under grant agreements No 654109 and 739530

Activities

Management and coordination

- Monthly management meetings
- Annual CiGas Community workshop
- Participation in RIComm and CF-Meetings; other working groups

Links with associated communities

- Standards Committee CEN WG13: „Ambient air - Ozone precursors and benzene”
- TOARII
- WMO-GAW: WCC for VOC
- MetClimVOC
- CAMS-21a-2nd phase; RI-URBANS; ATMO-ACCESS; EQUIPEX OBS4CLIM

Consultancy for NMHCs

Trainings for NMHCs measurement and data evaluation (planned 2023)

Ready for NMHC working standard and target gas checks





Certificate of Calibration

NPL PRIMARY REFERENCE MATERIAL

Cylinder Number: D933529

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CUSTOMER: Forschungszentrum Jülich GmbH
ADDRESS: Institut IEK-8: Troposphäre, Wilhelm-Johnen-Strasse, 52425 Jülich, Germany

CALIBRATION DATE: 08 December 2020


AMOUNT FRACTIONS:

Component	Amount fraction / (nmol/mol)	Component	Amount fraction / (nmol/mol)
Ethane	4.24 ± 0.13	2-methylpentane	4.39 ± 0.09
Ethene	4.15 ± 0.09	n-hexane	4.39 ± 0.09
Propane	4.18 ± 0.09	Isoprene	4.37 ± 0.09
Propene	4.15 ± 0.09	n-heptane	4.40 ± 0.09
2-methylpropane	4.26 ± 0.11	Benzene	3.72 ± 0.08
n-butane	4.22 ± 0.09	2,2,4-trimethylpentane	4.13 ± 0.09
Ethyne	4.37 ± 0.22	n-octane	4.14 ± 0.09
trans-but-2-ene	4.23 ± 0.09	Toluene	3.61 ± 0.10
But-1-ene	4.21 ± 0.09	Ethylbenzene	3.91 ± 0.10
cis-but-2-ene	4.22 ± 0.09	m-xylene + p-xylene	7.60 ± 0.20
2-methylbutane	4.16 ± 0.09	o-xylene	3.74 ± 0.10
n-pentane	4.18 ± 0.09	1,3,5-trimethylbenzene	3.80 ± 0.10
1,3-butadiene	4.27 ± 0.09	1,2,4-trimethylbenzene	3.83 ± 0.10
trans-pent-2-ene	4.20 ± 0.09	1,2,3-trimethylbenzene	3.80 ± 0.10
Pent-1-ene	4.27 ± 0.09	Nitrogen	Balance

The reported expanded uncertainties are based on standard uncertainties multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

METHODS: Preparation: gravimetry; Analysis: gas chromatography (FID)
TRACEABILITY: The values on this certificate are traceable to NPL Primary Standards
EXPIRY: Certificate valid for 5 years from the date of issue
PRESSURE: Fill pressure: 100 bar; Minimum utilisation pressure: 10 bar
STORAGE: No special precautions are required
HANDLING: Refer to ISO 16664
OUTLET: DIN 477 No. 1 valve
INTENDED USE: Calibration standard

Reference: 2021030009-1 **Date of issue:** 22 March 2021

Signed:  (Authorised Signatory)

Name: Dr D R Worton (on behalf of NPLML)

Checked by: 

Page 1 of 1



4002



Certificate of Calibration

NPL PRIMARY REFERENCE MATERIAL

Cylinder Number: D933592

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CUSTOMER: Forschungszentrum Jülich GmbH
ADDRESS: Institut IEK-8: Troposphäre, Wilhelm-Johnen-Strasse, 52425 Jülich, Germany

CALIBRATION DATE: 05 May 2021

AMOUNT FRACTIONS:

Component	Amount fraction / (nmol/mol)
Toluene	3.93 ± 0.12
(+/-)- α -pinene	4.17 ± 0.21
(+)-3-carene	3.93 ± 0.20
R-(+)-limonene	3.81 ± 0.12
1,8-cineole	4.08 ± 0.21
Nitrogen	Balance

The reported expanded uncertainties are based on standard uncertainties multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95 %. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

METHODS: Preparation: gravimetry; Analysis: gas chromatography (FID)
TRACEABILITY: The values on this certificate are traceable to NPL Primary Standards
EXPIRY: Certificate valid for 1 year from the date of issue
PRESSURE: Fill pressure: 100 bar; Minimum utilisation pressure: 10 bar
STORAGE: No special precautions are required
HANDLING: Refer to ISO 16664
OUTLET: DIN 477 No. 1 valve
INTENDED USE: Calibration standard

Reference: 2021030009-2

Date of issue: 07 May 2021

Signed:  (Authorised Signatory)


Name: Dr P J Brewer (on behalf of NPLML)

Checked by: 

Page 1 of 1

Measurement performance monitoring

Preparation for
round-robins



Measurement guidelines



HOME SCIENCE & TECHNOLOGY NATIONAL FACILITIES PARTNERS & USERS ANNOUNCEMENTS & RESOURCES

Publications & documents

Standard Operation Procedures

For NO_x and VOC can be found here [SOP for NO_x and VOC](#)




Deliverable 3.17. Updated Measurement Guideline for NO_x and VOCs

Stefan Reimann (EMPA), Robert Wegener (FZJ), Anja Claude (DWD), Stephan Sauvage (CNRS)

Work package no	WP3
Deliverable no.	D3.17. Updated Measurement Guideline for NO _x and VOCs
Lead beneficiary	EMPA
Deliverable type	<input checked="" type="checkbox"/> R (Document, report) <input type="checkbox"/> DEC (Websites, patent fillings, videos, etc.) <input type="checkbox"/> OTHER: please specify

op 2023; Online – April, 17th - 19th 2023



Data review




Empa

Materials Science and Technology




@VOC@



footprints

sabine.eckhardt@nilu.no

4days CV0001G Footprint regional FIRST -50 PREV NEXT +50 LAST 2/306 : 2019-03-02
SUBMIT



Activities

Labelling

- Develop and establish workflows, communication and management
- Contribution to the assessment for NMHCs
- Reporting for labelling step 1a


Preparation for auditing pilote NFs

- Provision of gas standards
- On site audit

s-b-s-Inter comparisions

Scientific developments

- Method developments and QA procedures for CIMS & TD-GC
- Low cost sensor applications
- Airborne applications of CIMS



Implementation Status (FZJ)



VOCUS



TD-GC-FID/MS
(Gerstel TDS-G; CIS)



PTR-TOF-MS
8000



Liquid
Calibration Unit
Certified
laboratory
standards (NPL)



Diffusion sources



Implementation Status (FZJ)



System for automated
Canister and sample tube
analysis connected to TD-
GC-FID/MS (Markes,
Agilent)



On-line TD-GC-FID/MS
(Markes, Agilent)
TD-GC-MS (Markes,
Agilent)



Gas mixing and continuous
calibration gas generation



Calibration flow tube
(CACTUS)

Implementation Plans (FZJ)

Upcoming:

- Gas mixing & cylinder filling → Working standard provision
- Target gas filling → Compressor
- PTR-TOF-MS (FUSION)
- HCHO TILDAS Instrument
- Hands-on training center + s-b-s-intercomparison facility connected to SAPHIR and JULIAC

TILDAS Compact Single Laser Formaldehyde Analyzer

Exceptional HCHO accuracy and precision in a compact, rugged package.



FUSION
PTR



Implementation status (Overview)

Activity	NMHCs	Implementation year
Consultancy	anthropogenic	Available
	biogenic	2024
Training	anthropogenic	2023
	biogenic	2024
Measurement Guidelines	anthropogenic	Available
	(update)	2023
	biogenic	2024
Scale available at NFs	both	Available
Working standard checks	anthropogenic	Available
	biogenic	2024
Measurement performance monitoring	anthropogenic	2023
	biogenic	2025
Data review	anthropogenic	Available
	biogenic	2024
Analytical instrument compatibility test	both	2025
Audits	anthropogenic	2023
	biogenic	2024
ACTRIS approved new technologies	both	2025



FZJN Implementation



Challenges

NO:

Traceable gas standard available from CCL

NO level span three orders of magnitude (!) at different ACTRIS sites

Most Sensitive CL reference instrument will be purchased
(April 2023)



FZJN Implementation

Challenges

NO₂

Traceable gas standard available from CCL

NO level CLD Measured after conversion into NO is prone to interferences

Setup of:



Spectroscopic (direct!) reference method for direct NO₂ measurement onsite (mobile !)

DOAS instruments operating.

First deployment at campaign at Hohenpeissenberg shows excellent performance



DOAS spectrometers



Comparison of FZJN DOAS with TDL spectrometer at Hohenpeissenberg

FZJN Implementation

Challenges


NO₂

No traceable NO₂ gas standard available

NO₂ for calibration is produced from NO and ozone

Setup of:

Mobile calibration stand with a gas phase titration and humidification device and NO / NO₂ and CO₂ measurements for checking NF devices by measurements on site. Calibration stand set up., will be replicated in 2023.



Mobile Calibration Unit

The rack will also contain ozone and water measurements allow to check for interferences at the NFs

FZJN Implementation

SAPHIR



SAPHIR chamber with JULIAC tower

All instruments are routinely operated in parallel at ambient concentrations at SAPHIR to check for potential problems, such as interference.

Here, different techniques (CAPS, TDL, DOAD, CL) are compared to each other.

A new lab building will be setup close to the chamber to host NO_x and VOC during intercomparison



Implementation Status (IMT)



IMT Nord Europe
École Mines-Télécom
IMT-Université de Lille



Available:


- PTR-QiToF-MS Ionicon
- PTRMS Kore
- TD-GC FID/FID; TD-GC-FID/MS (offline & online)
- Liquid Calibration Unit; Gas Calibration Unit
- Permeation system
- Target gas cylinder filling system
- Multi-gas generation systems & intercomparison platform

Upcoming:

- VOCUS 2023-2024


+ Certified laboratory standards (NPL, NIST, upcoming VSL)



- Monitoring of reactive gases @MOHp since 1994 (GAW)
→ Monitoring of VOCs @MOHp since 1998 (GAW+EMEP)
- In ACTRIS since 2011: involved in the development of QA/QC measures, measurement guidelines and organisation of intercomparison campaigns (side-by-side and RR)



Monitoring at HPB:

- **VOC (>50 NMHC, BVOC, OVOC; 12h intervals; 3 GC systems):**
 - TD-GC-FID (C2-C8 NMHCs),
 - TD-GC-MC/FID (biogenic VOCs),
 - TD-GC-FID/FID(/MS) (NMHCs+OVOCs)
 - PTR-TOF-MS (2023)
- CO (cavity + fluorescence)
- O₃ (UV-absorption)
- NO (CLD)
- NO₂ (CLD-PLC/BLC + CAPS)
- NO_y (gold converter CLD + TD-CAPS)
- SO₂ (fluorescence)
- CO₂, CH₄, N₂O (ICOS – cavity absorption)
- OH, H₂SO₄, K(OH), RO_x (CIMS)





Measurement of NMHCs by GC

On-Line sampling

- Inlet location >2 m above the building and >5 m above ground
- Sample path including inlet line and filters inert to NMHCs → surface passivated steel, glass
- Controlled sampling flow
- Removal of water, ozone, carbon dioxides and particles
- Sample pre-concentration (low temperature adsorption)
- Transfer of the analyte to GC (by carrier gas flow, high temp., potentially 2nd focussing trap)
- FID and/or MS detection

Off-line sampling (additional requirements with respect to on-line sampling)

- Passivated or electropolished stainless steel canisters
- Glass flasks
- Storage times max. 30 days

Data quality objectives for NMHCs measured by GC

Table 3 Data quality objectives (DQOs) for the measurements of NMHCs in whole air compressed test gases (inter-laboratory compatibility) expressed as the expanded combined uncertainty ($k=2$) and the repeatability ($k=1$; standard deviation). The basic station performance requirements correspond to the former and weaker DQOs of GAW Report 171 (2006).

	GAW basic performance expanded combined uncertainty	GAW basic performance repeatability	ACTRIS target performance expanded combined uncertainty	ACTRIS target performance repeatability
Alkanes	10%	5%	5%	2%
alkenes incl. isoprene	20%	10%	5%	2%
Alkynes	15%	5%	5%	2%
Aromatics	15%	10%	5%	2%
mole fraction ⁽¹⁾	10/15/20			
<100 pmol/mol	pmol/mol	5/10 pmol/mol	5 pmol/mol	2 pmol/mol

⁽¹⁾For mole fractions below 100 pmol/mol, the DQO are expressed in pmol/mol, reference is the above stated relative value at 100pmol/mol e.g. for alkanes basic performance 10 pmol/mol.



NMHC-Gas Standard Requirements

The Central Calibration Laboratory (CCL) maintains the primary standard that defines the calibration scale. → NPL (UK)

Standard requirements for ACTRIS-NFs:

1. (secondary) **Laboratory standard:** multi-component standard (synthetic mixture), produced and certified by the CCL.
2. (tertiary) **Working standards:** Cover most (ideally all) components measured and are used for regular calibration. WS can be other-certified or custom-made synthetic mixtures, or compressed whole air, calibrated by CiGas.
3. A **target gas mixture:** Compressed whole air / synthetic mixture calibrated by CiGas.

The target gas is used to check the assigned values of the calibration mixtures and the calibration process itself, and is treated as an air sample with unknown amount fraction. Monitoring the target gas concentrations yields information about the performance of the instrument, drifts of the laboratory standard, and potential instrument problems.

Recommended frequencies for standard, blank and target gas measurements

System	Lab. Standard	Working Standard	Blank	Target gas	Stand. Addition
GC-FID	2/year*	2/month*	1/week	1/month*	1/year*
GC-MS	2/year*	Every 2-4 samples	1/week	1/month*	1/year*

*) 3-5 replicates